



TECHNICAL REPORT

GASSARMATUR A/S

OXYGEN SHOCK TESTING OF THREE OFF OXYGEN
PIGTAILS.


REPORT No. 2008-3049

REVISION No. 01

DET NORSKE VERITAS



TECHNICAL REPORT

Date of first issue: 24 January 2008	Project No.: 71523185
Approved by: Stein Fredheim Head of Section 	Organisational unit: Structural Integrity and Laboratories
Client: Gassarmatur A/S	Client ref.:

DET NORSKE VERITAS
REGION NORGE AS REGION
NORDIC COUNTRIES

Veritasveien 1,
N-1322 HØVIK, Norway
Tel: +47 67 57 99 00
Fax: +47 67 57 99 11
http://www.dnv.com
Org.No: NO 959 627 606 MVA



Summary:

Oxygen shock testing of three off oxygen Pigtails, drawing No. 150607, has been carried out.

The test of the Pigtails was based on ISO 2503: 1998, Clause 5.11.3. Ignition test (oxygen pressure regulators).

No visible damage or sign of burning were observed as a result of the testing and all three Pigtails passed the test.

The test was carried out in the Det Norske Veritas' Laboratory at Høvik, Norway, 23 January 2008.

Report No.: 2008-3049	Subject Group: E9	
Report title: Oxygen shock testing of three off oxygen Pigtails.		
Work carried out by: Jan Lasse Moe 		
Work verified by: Odd Sund 		
Date of this revision:	Rev. No.:	Number of pages:
	01	6

Indexing terms

Oxygen shock testing
Pigtail

- No distribution without permission from the Client or responsible organisational unit
- Limited distribution within Det Norske Veritas
- Unrestricted distribution